

## Patient Identification Card

This person is implanted with a medical hearing device manufactured by Cochlear Limited. A metal device is implanted under the skin behind the ear, and external equipment is worn behind the ear or on the body (e.g. in a pocket). Without the equipment, this person cannot hear.

The system may activate airport security metal detectors.



The patient must remove all external components of their implant system before entering a room where an MRI scanner is located. This person may be safely scanned with MRI only under very specific conditions for their implant model. Scanning under different conditions may result in severe patient injury or device malfunction. For full MRI safety information refer to the information supplied with the system, or contact your regional Cochlear office – contact numbers available on reverse side of card.

*Hear now. And always*



First name: \_\_\_\_\_ Last name: \_\_\_\_\_

Address: \_\_\_\_\_

\_\_\_\_\_ Country: \_\_\_\_\_ Tel: \_\_\_\_\_

Emergency contact: \_\_\_\_\_ Tel: \_\_\_\_\_

Implant serial no.: \_\_\_\_\_ Model: \_\_\_\_\_ Ear: \_\_\_\_\_

Implant serial no.: \_\_\_\_\_ Model: \_\_\_\_\_ Ear: \_\_\_\_\_



**Cochlear Ltd** (ABN 96 002 618 073) 1 University Avenue, Macquarie University, NSW 2109, Australia

Tel: +61 2 9428 6555 Fax: +61 2 9428 6352

**Cochlear Ltd** (ABN 96 002 618 073) 14 Mars Road, Lane Cove, NSW 2066, Australia

Tel: +61 2 9428 6555 Fax: +61 2 9428 6352

**www.cochlear.com**

**EC/REP Cochlear Deutschland GmbH & Co. KG**  
Karl-Wiechert-Allee 76A, 30625 Hannover, Germany

Tel: +49 511 542 770 Fax: +49 511 542 7770

**Cochlear Americas** 13059 E Peakview Avenue,  
Centennial, CO 80111, USA

Tel: +1 303 790 9010 Fax: +1 303 792 9025



P1109247 D1109237-V2 JUN19

D1109237-V2