

Cochlear™ Osia®

Magnetic Resonance Imaging (MRI) Guidelines

Hear now. And always



Cochlear®

About this document

This guide applies to the Cochlear™ Osia® OSI200 Implant. It is intended for:

- Specialised health care professionals who prepare and perform MR scans
- Physicians who refer a Cochlear Osia implant recipient for an MR scan
- Cochlear Osia implant recipients and/or their carers.

This guide provides information about the safe application of an MR scan on Cochlear Osia OSI200 Implant recipients.

MR scans performed under different conditions than those presented in this guide may result in severe patient injury or device malfunction.

Due to the risks associated with using MRI with an implanted medical device, it is important to read, understand, and comply with these instructions to prevent potential harm to the patient and/or device malfunction.

This guide should be read in conjunction with the relevant documents that accompany a Cochlear Osia OSI200 Implant, such as the *Physician's Guide* and *Important information for Osia implant recipients*.

For more information, contact Cochlear by calling your regional Cochlear office – contact numbers are available on the back cover of this guide or visit www.cochlear.com/warnings.

Symbols used in this document



NOTE

Important information or advice.



CAUTION (NO HARM)

Special care to be taken to ensure safety and effectiveness.
Could cause damage to equipment.



WARNING (HARMFUL)

Potential safety hazards and serious adverse reactions.
Could cause harm to person.

Contents

About this document.....	2
Symbols used in this document.....	2
MRI safety information	4
MRI safety information for Osia OSI200 Implant with BI300 Implant	6
1.5 T scans.....	6
3 T scans	7
Preparation prior to an MRI examination.....	8
Cooperation between specialists	8
Considerations for implant magnet removal	9
Considerations for conducting an MRI examination	10
Prerequisites.....	10
Considerations for referring physicians	11
Risks associated with MRI and Cochlear Osia implants.....	12

MRI safety information

In order to determine if a patient may receive an MR scan, you must first identify the patient's implant model.

After you have identified the implant model, locate the MRI safety information for that specific implant model.



All external components of the Cochlear Osia System (e.g. sound processors, and related accessories) are MR Unsafe. The patient must remove all external components of their system before entering a room where an MRI scanner is located.

Identifying the Cochlear Osia implant

The implant model can be found on the patient's Cochlear patient identification card.

If the patient does not have their patient identification card with them, the implant type and model can be identified without surgical intervention using X-ray or the Cochlear Osia fitting software.

X-ray information

Cochlear Osia OSI200 Implants are made of metal and implanted under the skin behind the ear. Using an X-ray, the implant can be identified by its shape and the shape of the actuator unit.

Use the Fig. 1 and the Fig. 2 to assist with identifying the differences between Cochlear Osia implant models when using an X-ray.




Fig. 1 Approximate location of the OSI200 Implant



Fig. 2 OSI200 Implant (P1170466)

MRI safety information for Osia OSI200 Implant with BI300 Implant



Non-clinical testing has demonstrated that the Osia OSI200 Implant, in combination with the BI300 Implant, are MRI Conditional. Patients can be scanned at 1.5 T and 3 T only if the magnet has been removed. A patient with these devices can be safely scanned in an MR system meeting the following conditions:

1.5 T scans

- Surgically remove the implant magnet before MR scans at 1.5 T. See OSI200 Implant Physician's Guide for additional information.
- Remove the sound processor before entering the MRI scan room. The sound processor is MR Unsafe.
- Static magnetic field of 1.5 T with the implant magnet surgically removed.
- Maximum spatial field gradient of 2000 gauss/cm (20 T/m).
- When using a transmit/receive head coil, a maximum MR system reported, head averaged specific absorption rate (SAR) of 3.2 W/kg.
- When using a transmit body coil, a maximum MR system reported, whole body averaged specific absorption rate (SAR) of 2 W/kg.

In non-clinical testing, the maximum image artefact caused by the OSI200 Implant when imaged with a gradient echo pulse sequence scan in the axial plane is as follows:

Non-Magnetic Plug	No Magnet
6 cm (2.4 in)	6 cm (2.3 in)

Table 1 Maximum image artefact from centre at 1.5 T (gradient echo sequence).

3 T scans

- Surgically remove the implant magnet before MR scans at 3 T. See OSI200 Implant Physician's Guide for additional information.
- Remove the sound processor before entering the MRI scan room. The sound processor is MR Unsafe.
- Static magnetic field of 3 T with the implant magnet surgically removed.
- Maximum spatial field gradient of 2000 gauss/cm (20 T/m).
- When using a transmit/receive head coil, a maximum MR system reported, head averaged specific absorption rate (SAR) of 3.2 W/kg.
- When using a transmit body coil, a maximum MR system reported, whole body averaged specific absorption rate (SAR) of 2 W/kg.

Scans must be performed in circular polarization mode.

In non-clinical testing, the maximum image artefact caused by the OSI200 Implant when imaged with a gradient echo pulse sequence scan in the axial plane is as follows:

Non-Magnetic Plug	No Magnet
5 cm (2 in)	5 cm (2 in)

Table 2 Maximum image artefact from centre at 3 T (gradient echo sequence).

Preparation prior to an MRI examination

Cooperation between specialists

Preparing for and conducting an MRI examination for implant recipients requires cooperation between a specialist for the device and/or Osia implant physician, referring physician and radiologist/MR technologist.

Cochlear Osia implant device specialist

Knows the implant type and where to find the correct MR parameters for the implant.

Referring physician

Knows the location of the MR scan and diagnostic information required.

Cochlear Osia implant physician

Surgically removes the implant magnet and replaces it with a new sterile replacement implant magnet.

Radiologist/MR technologist

Sets up the MR scan using the correct MR parameters and counsels the implant recipient during the MRI examination.

Considerations for implant magnet removal

Close coordination is required between the specialists to perform the implant magnet removal, MR scan, and subsequent implant magnet replacement. For details on implant magnet removal, please refer to the *OSI200 Implant Physician's Guide* supplied with the system.

For implant recipients requiring multiple MRI examinations over a period of time, the implant magnet is removed and replaced with a sterile non-magnetic plug. In the magnet's absence, the non-magnetic plug prevents fibrous tissue growing into the implant recess. Such growth would make implant magnet replacement difficult.

While the implant magnet is removed, the recipient must wear a retainer disc to hold their sound processor coil in place. Retainer discs are available from Cochlear.

When there is no further need for MRI examinations, the non-magnetic plug is removed and replaced by a new sterile replacement implant magnet.

The sterile non-magnetic plug and sterile replacement magnet are supplied separately in sterile packs. Both are single-use items.

Considerations for conducting an MRI examination

These guidelines are specific to Cochlear Osia implants and supplement other MRI examination considerations specified by the MRI machine manufacturer or protocols at the MRI facility.

Prerequisites

The following additional conditions must be met:

- The implant model has been identified.
- The implant magnet has been surgically removed when the referring physician has prescribed that the MR scan be performed with the implant magnet removed.

Patient positioning

The patient should be positioned prior to entering the MRI machine. The patient should be placed in the supine position (lying flat on back, face upward), with their head aligned with the bore axis of the MRI machine.

The patient should be advised to lie as still as possible and to not move their head during the MR scan.



- Ensure that the patient does not move more than 15 degrees (15°) from the centreline (Z-axis) of the bore during the MR scan.
- Failure to position the patient correctly prior to the MR scan may result in increased torque on the implant and cause pain.



Explain to the patient that they may perceive sounds during the MR scan. The sound levels are not hazardous.

Perform the MR scan

The MR scan must be performed using the MRI safety information identified for the patient's implant model.

Performing an MR Scan on other body locations

When an implant recipient requires an MRI on a location of their body away from the implant site, you must still follow the MRI safety information for the recipient's implant model. See Implant model identification and related *MRI safety information* on page 6.

Considerations for referring physicians

If you are a physician referring a Cochlear Osia implant recipient for an MR scan, it is essential that you consider the following:

- Understand and inform the patient of the risks associated with MRI. See *Risks associated with MRI and Cochlear Osia implants* on page 17.
- Understand the conditions for an MR scan and ensure that there is a clear indication for the MRI examination.
- Identify if the patient has any other medical device implants, active or abandoned. If another implant is present, verify MRI compatibility before conducting an MRI procedure. If MRI guidelines for specific devices are not followed potential risks include: movement or damage to the device, weakening of the implant magnet and uncomfortable sensation or skin/tissue trauma for the patient.
- The Cochlear Osia implant will create shadowing on the MR image in the vicinity of the implant, resulting in a loss of diagnostic information.
- See *MRI safety information for Osia OS1200 Implant with B1300 Implant* on page 6.

Consider the following:

- Timing of the implant surgery and MRI exposure.
- Age and general health of the implant recipient and time to recover from the implant magnet surgery or potential trauma.
- Existing or potential for tissue scarring in the location of the implant magnet.
- When the implant magnet needs to be removed, refer the patient to an appropriate physician to arrange for the magnet to be removed before the MR scan.

Risks associated with MRI and Cochlear Osia implants

The potential risks of performing MRI examinations on patients with Cochlear Osia implants include:

Damage to the device

MRI exposure beyond the values contained in these guidelines may cause damage to the device.

Uncomfortable sensation

MRI exposure beyond the values contained in these guidelines may result in the patient perceiving sound or noise and/or pain.

Implant heating












Use the recommended SAR values contained in these guidelines to ensure the implant does not heat beyond safe levels.

Image artefact

The Cochlear Osia implant will create shadowing on the MR image in the vicinity of the implant, resulting in a loss of diagnostic information.

Labelling symbols

The following symbols may appear on the product, the components and/or the packaging.

	Refer to instruction manual		Keep dry
	Specific warnings or precautions associated with the device, which are not otherwise found on the label		Do not re-use
	Manufacturer		Do not use if package is damaged
	Date of manufacture	Rx Only	By prescription
	Catalogue number		MR Conditional
	Authorised representative in the European Community		MR Unsafe

Hear now. And always



- Cochlear Ltd** (ABN 96 002 618 073) 1 University Avenue, Macquarie University, NSW 2109, Australia
Tel: +61 2 9428 6555 Fax: +61 2 9428 6352
- Cochlear Ltd** (ABN 96 002 618 073) 14 Mars Road, Lane Cove, NSW 2066, Australia
Tel: +61 2 9428 6555 Fax: +61 2 9428 6352
- CECOP Cochlear Deutschland GmbH & Co. KG** Karl-Wiechert-Allee 76A, 30625 Hannover, Germany
Tel: +49 511 542 770 Fax: +49 511 542 770
- Cochlear Americas** 13059 E Peakview Avenue, Centennial, CO 80111, USA
Tel: +1 303 790 9010 Fax: +1 303 792 9025
- Cochlear Canada Inc** 2500-120 Adelaide Street West, Toronto, ON M5H 1T1, Canada
Tel: +1 416 972 5082 Fax: +1 416 972 5083
- Cochlear AG** EMEA Headquarters, Peter Merian-Weg 4, 4052 Basel, Switzerland
Tel: +41 61 205 8204 Fax: +41 61 205 8205
- Cochlear Europe Ltd** 6 Dashwood Lang Road, Bourne Business Park, Addlestone, Surrey KT15 2HJ, United Kingdom
Tel: +44 1932 26 3400 Fax: +44 1932 26 3426
- Cochlear Benelux NV** Schaliënhoedreef 20 i, B-2800 Mechelen, Belgium
Tel: +32 15 79 55 11 Fax: +32 15 79 55 70
- Cochlear France S.A.S.** 135 Route de Saint-Simon, 31035 Toulouse, France
Tel: +33 5 34 63 85 85 (International) or 0805 200 016 (National) Fax: +33 5 34 63 85 80
- Cochlear Italia S.r.l.** Via Larga 33, 40138 Bologna, Italy
Tel: +39 051 601 53 11 Fax: +39 051 39 20 62
- Cochlear Nordic AB** Konstruktionsvägen 14, 435 33 Mölnlycke, Sweden
Tel: +46 31 335 14 61 Fax: +46 31 335 14 60
- Cochlear Tibbi Çihazlar ve Sağlık Hizmetleri Ltd. Şti.**
Çubuklu Mah. Boğaziçi Cad., Boğaziçi Plaza No: 6/1, Kavaçık, TR-34805 Beykoz-İstanbul, Turkey
Tel: +90 216 538 5900 Fax: +90 216 538 5919
- Cochlear (HK) Limited** Room 1404-1406, 14/F, Leighton Centre, 77 Leighton Road, Causeway Bay, Hong Kong
Tel: +852 2530 5773 Fax: +852 2530 5183
- Cochlear Korea Ltd** 1st floor, Cheongwon Building 33, Teheran-ro 8 gil, Gangnam-gu, Seoul, Korea
Tel: +82 2 533 4450 Fax: +82 2 533 8408
- Cochlear Medical Device (Beijing) Co Ltd**
Unit 2608-2617, 26th Floor, No.9 Building, No.91 Jianguo Road, Chaoyang District, Beijing 100022, P.R. China
Tel: +86 10 5909 7800 Fax: +86 10 5909 7900
- Cochlear Medical Device Company India Pvt. Ltd.**
Ground Floor, Platina Building, Plot No C-59, G-Block, Bandra Kurla Complex, Bandra (E), Mumbai – 400 051, India
Tel: +91 22 6112 1111 Fax: +91 22 6112 1100
- 株式会社日本コクレア (Nihon Cochlear Co Ltd)** 〒113-0033 東京都文京区本郷2-3-7 お茶の水元町ビル
Tel: +81 3 3817 0241 Fax: +81 3 3817 0245
- Cochlear Middle East FZ-LLC**
Dubai Healthcare City, Al Razi Building 64, Block A, Ground Floor, Offices IR1 and IR2, Dubai, United Arab Emirates
Tel: +971 4 818 4400 Fax: +971 4 361 8925
- Cochlear Latinoamérica S.A.**
International Business Park, Building 3835, Office 403, Panama Pacifico, Panama
Tel: +507 830 6220 Fax: +507 830 6218
- Cochlear NZ Limited**
Level 4, Takapuna Towers, 19-21 Como St, Takapuna, Auckland 0622, New Zealand
Tel: +64 9 914 1983 Fax: 0800 886 036

www.cochlear.com

This content is meant for professional use. If you are a consumer, please seek advice from your medical practitioner or health professional about treatments for hearing loss. They will be able to advise you on a suitable solution for your hearing loss condition. All products should be used only as directed by your medical practitioner or health professional. Not all products are available in all countries. Please contact your local Cochlear representative.

Cochlear, 科利耳, コクレア, 코클리어, Hear now. And always, Osia, SmartSound, the elliptical logo, and marks bearing an ® or ™ symbol, are either trademarks or registered trademarks of Cochlear Bone Anchored Solutions AB or Cochlear Limited. © Cochlear Limited 2019. All rights reserved. NOV19.

